

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 18 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130190</u> ✓	
A	Name <u>J. Swartz</u>	ID# <u>4314</u> Date <u>3/30/09</u> ✓	
	Agency <u>SCDL</u>	Phone # <u>269-5592</u>	
	Instrument Location <u>SCDL</u>	✓	
B	Alco S/N <u>SC07108</u>	Target Value <u>.080</u> ✓ High Pressure <u>850 psi</u>	
	Alco Test Values	<u>.081</u> ✓ 1 <sup>st</sup> Alco	<u>.086</u> ✓ 2 <sup>nd</sup> Alco
	Signature <u>[Signature]</u>		
(OVER)			

BMB  
9/9/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 29<sup>th</sup> day of Sept, 2009.

[Signature]  
Carolyn M. Noland

(Notary Seal Stamp)

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130190

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

### Breath Test Record

Serial Number 130190 Model DataMaster  
Location ANC1 Anchorage Police Dept.

### Subject Information

Last Name Verification Of Cal O.L. # L  
First Name L

### Operator Information

Operator Name Swartz/J Case   
Operator # 4314 Department AAA1

### Test Results

Alco Target .080 Alco Serial SCOTT08  
Nominal 0.078 Barometer 29.31 Test Type V  
Dry Gas 1 0.081 Time: 07:13:47 Test Date 03/30/2009  
Subject Alcohol 0.000 Time: 07:15:03 Test Time 07:12:32  
Dry Gas 2 0.080 Time: 07:16:02 Report Y

## DIAGNOSTIC CHECK

### Diagnostic Test Record

Serial Number 130190 Model DataMaster  
Location ANC1 Anchorage Police Dept.

### Diagnostic Test Results

Test Date 03/30/2009 Computer Okay  
Test Time 07:10:55 Program Okay  
Software Date 01/08/2009  
Sample Chamber (C) 49  
Breath Tube (C) 40  
Barometer (in) 29.31  
Flow Detector Okay  
High Speed Pump Okay  
Detector Okay  
Filter Okay  
Quartz Okay  
Calibration Okay